



European
Commission

JRC VALIDATED METHODS, REFERENCE
METHODS AND MEASUREMENTS REPORTS

Report on the Validation of a DNA Extraction Method for Soybean Seeds

Corrected version 1 – 10 November 2014

Replace version – 18 February 2008

2014



Joint
Research
Centre

EURL
European Union Reference Laboratory
for GM Food & Feed

This document replaces the document CRLVL05/06XP with the same title published on the EU-RL Website. The correction made in the new document is a correction of a table containing Comparison of extrapolated Ct values versus measured Ct values.

European Commission

Joint Research Centre
Institute for Health and Consumer Protection

Contact information

Molecular Biology and Genomics Unit
Address: Joint Research Centre, Via Enrico Fermi 2749, TP 201, 21027 Ispra (VA), Italy
E-mail: eurl-gmff@jrc.ec.europa.eu
Tel.: +39 0332 78 5165
Fax: +39 0332 78 9333

<https://ec.europa.eu/jrc>

This publication is a Validated Methods, Reference Methods and Measurements Report by the Joint Research Centre of the European Commission.

Legal Notice

This publication is a Validated Methods, Reference Methods and Measurements Report by the Joint Research Centre, the European Commission's in-house science service. It aims to provide evidence-based scientific support to the European policy-making process. The scientific output expressed does not imply a policy position of the European Commission. Neither the European Commission nor any person acting on behalf of the Commission is responsible for the use which might be made of this publication.

JRC94638

© European Union, 2014

Reproduction is authorised provided the source is acknowledged.

Printed in Italy



EUROPEAN COMMISSION
JOINT RESEARCH CENTRE

Institute for Health and Consumer Protection
Molecular Biology and Genomics Unit



Report on the Validation of a DNA Extraction Method for Soybean Seeds

Report on the Validation of a DNA Extraction Method for Soybean Seeds

18 February 2008

Corrected version 1 - 10 November 2014

Method development and single laboratory validation:

Monsanto Company

Method testing and confirmation:

European Union Reference Laboratory for GM Food and Feed (EU-RL GMFF)

Quality assurance

The EU-RL GMFF is ISO 17025:2005 accredited [certificate number: ACCREDIA 1172, (Flexible Scope for DNA extraction and qualitative/quantitative PCR) - Accredited tests are available at http://www.accredia.it/accredia_labsearch.jsp?ID_LINK=293&area=7].

The original version of the document containing evidence of internal checks and authorisation for publication is archived within the EU-RL GMFF quality system.

The EU-RL GMFF is also ISO 17043:2010 accredited (proficiency test provider) and applies the corresponding procedures and processes for the management of ring trials during the method validation.

The EU-RL GMFF conducts its activities under the certification ISO 9001:2008 of the Institute for Health and Consumer Protection IHCP provided by CERMET.

Corrections from previous version:

Page 10/11 Table 3. Comparison of extrapolated Ct values versus measured Ct values (amplification of soybean lectin gene, lec)
*Values in the column ΔCt^{**} have been corrected*

Address of contact laboratory:

European Commission, Joint Research Centre
Institute for Health and Consumer Protection (IHCP)
Biotechnology and GMOs Unit – Community Reference Laboratory for GM Food and Feed
Via Fermi 2749, 21027 Ispra (VA) - Italy

Content

1. INTRODUCTION.....	4
2. MATERIALS (EQUIPMENT/CHEMICALS/PLASTICWARE).....	4
3. DESCRIPTION OF THE METHODS	6
4. TESTING OF THE DNA EXTRACTION METHOD BY THE COMMUNITY REFERENCE LABORATORY FOR GM FOOD AND FEED	8
5. CONCLUSION	11
6. QUALITY ASSURANCE.....	11
7. REFERENCES.....	11

1. Introduction

This report describes the validation of a DNA extraction protocol derived from the publicly available "CTAB" method ⁽¹⁾ and its applicability on the samples of food and feed provided by the applicant. This protocol can be used for the extraction of DNA from soybean seeds.

The purpose of the DNA extraction method described is to provide DNA with purity suitable for real-time PCR based detection methods.

This protocol is recommended to be executed by skilled laboratory personnel as the procedures comprise the use of hazardous chemicals and materials. It is strongly advised to take particular notice of products safety recommendations and guidelines.

2. Materials (Equipment/Chemicals/Plasticware)

2.1. Equipment

The following equipment is used in the DNA extraction procedure described (equivalents may be used):

1. Centrifuge (Beckman Coulter Avanti J-251)
2. Shaker (LabLine Enviro 3527)
3. Thermometer (VWR Cat. No. 61222-504)
4. Vacufuge (Eppendorf 5301 22 82 010-9)
5. Water bath (Precision Cat. No. 51220046)
6. Microcentrifuge (Any appropriate model)

2.2. Chemicals

The following chemicals are used in the DNA extraction procedure described (equivalents may be used):

1. 24:1 chloroform:isoamyl alcohol (Sigma Cat. No. C-0549)
2. 25:24:1 phenol:chloroform:isoamyl alcohol (Sigma Cat. No. P-3803)
3. Ammonium acetate 7.5 M (Sigma Cat. No. A-2706)
4. CTAB (Sigma Cat. No. H-6269)
5. 0.5 M EDTA, pH 8.0 (GibcoBRL Cat. no. 15575-038)
6. 100% ethanol (AAPER)
7. NaCl (Sigma Cat. No. S-5150)
8. 2-mercaptoethanol (Bio-Rad Cat. no. 161-0710)
9. RNase A (Roche Cat. No. 10 109 196 001)

10. Isopropanol (EM Science Cat. No. PX1835-9)
11. 1 M Tris HCl pH 8.0 (Sigma Cat. No. T-3038)
12. Proteinase K (Roche Cat. No. 03 115 836 001)
13. Polyethylene Glycol (MW 8000) (Sigma Cat. No. P2139)

2.3. Solutions

The following buffers and solutions are used in the DNA extraction procedure described:

1. CTAB Extraction Buffer (2%) (store at room temperature)
 - 2% w/v CTAB
 - 100 mM Tris HCl pH 8.0
 - 20 mM EDTA pH 8.0
 - 1.4 M NaCl
2. Tris-EDTA buffer (TE 1X) (store at room temperature)
 - 10 mM Tris HCl pH 8.0
 - 1 mM EDTA pH 8.0
3. Proteinase K (10 mg/ml) (store at -20 °C)
4. RNase A (10 mg/ml) (store at -20 °C)
5. PEG Precipitation Buffer (20% w/v) (store at room temperature)
6. Ethanol (70% v/v) (store at room temperature)
7. Ethanol (80% v/v) (store at room temperature)

2.4. Plasticware

1. 50 ml conical tubes (Corning Cat. No. 430290)
2. 13 ml Sarstedt tubes (Sarstedt Cat. No. 60.540)
3. 1.5 ml microcentrifuge tubes
4. filter tips

Note: All plasticware has to be sterile and free of DNases, RNases and nucleic acids.

2.5. Precautions

- Phenol, chloroform, isoamyl alcohol, and isopropanol are hazardous chemicals; therefore, all manipulations have to be performed according to safety guidelines, under fume hood.
- It is recommended to use clean containers for Waring blenders for grinding the seed bulk samples.

- All tubes and pipette tips have to be discarded as biological hazardous material

2.6 Abbreviations:

EDTA	ethylenediaminetetraacetic acid
PCR	polymerase chain reaction
RNase A	ribonuclease A
TE	Tris EDTA
Tris	Tris(hydroxymethyl)aminomethane

3. Description of the methods

Sampling:

For sampling methods, it is referred to the technical guidance documents and protocols described in Commission Recommendation 2004/787/EC on technical guidance for sampling and detection of genetically modified organisms and material produced from genetically modified organisms as or in products in the context of Regulation (EC) N. 1830/2003.

Scope and applicability:

The method for DNA extraction described below is suitable for the isolation of genomic DNA from a wide variety of soybean tissues and derived matrices. However, validation data presented here are restricted to ground soybean seeds. Application of the method to other matrices may require adaptation and possible further specific validation.

Principle:

The basic principle of the DNA extraction consists of first releasing the DNA present in the matrix into aqueous solution and further purifying the DNA from PCR inhibitors. The present method starts with a lysis step (thermal lysis in the presence of Tris HCl, EDTA, CTAB and β -mercaptoethanol) followed by removal of contaminants such as lipophilic molecules and proteins by extraction with phenol and chloroform.

A DNA precipitate is then generated by using isopropanol. The pellet is dissolved in TE-buffer. Remaining inhibitors are removed by PEG precipitation and re-suspension in TE-buffer.

Tissues crushing procedure:

Tissues should be processed prior to extraction procedure. Possible methods of processing include a mortar and pestle with liquid nitrogen (leaf) or commercial blender (grain or seed).

Extraction of genomic DNA from soybean seed

1. Weight out 6 g of processed tissue into a 50 ml conical tube appropriate for centrifugation. Note: For unprocessed tissue, weighing may occur prior to processing as long as entire processed sample is transferred to the conical tube.
2. For each 6 g sample add 25 ml of a solution consisting of 24.25 ml, pre-warmed CTAB extraction buffer, 0.5 ml 2-mercaptoethanol (2-ME), and 0.25 ml of 10 mg/ml proteinase K for a final concentration of 2% (2-ME) and 100 µg/ml (proteinase K).
3. Incubate for 60 minutes at 55 °C. Cool briefly on bench (10 minutes)
4. Add 20 ml of phenol:chloroform:isoamyl alcohol (PCI, 25:24:1). Cap the tube and mix vigorously by vortex or inversion.
5. Centrifuge for 10 minutes at 13000 x g and 20-25 °C to separate the aqueous and organic phases. Transfer upper aqueous phase to a clean 50 ml conical tube.
6. Repeat extraction two times for a total of three extractions (step 4-5).
7. Transfer upper aqueous phase to a new tube and add approximately 2/3 volume of -20 °C isopropanol and gently invert the tube several times to mix.
8. To precipitate the DNA place the tube at -20 °C for 30 minutes. DNA may be stored as isopropanol precipitate at -20 °C for up to 1 year.
9. Centrifuge the tubes at approximately 13000 x g for 20 minutes at 4 °C to pellet the DNA
10. Re-dissolve the pellet in 4 ml of TE pH 8.0. Transfer to a 13-ml Sarstedt tube and add approximately 40 µl of 10 mg/ml RNase, then incubate at 37 °C for 30 minutes.
11. To extract the DNA add 4 ml of chloroform:isoamyl alcohol (CIA, 24:1). Centrifuge for 10 minutes at approximately 13000 x g at room temperature. Transfer the upper aqueous phase to a clean Sarstedt tube.
12. Repeat step 11 twice, then add half volume of 7.5 M ammonium acetate, gently mix by inversion/pipetting and add 2 volumes of 100% ethanol. Mix by inversion/pipetting and place at -20 °C for 30 minutes. DNA may be stored as ethanol precipitate at -20 °C for up to 1 year.
13. Centrifuge at 13000 x g for 20 minutes at 4 °C to pellet the DNA.
14. Rinse the DNA pellet twice with 70 % ethanol and remove residual ethanol by vacuum.
15. Re-suspend DNA in 1 ml TE, pH 8.0 and incubate at 65 °C for 1 hour with periodic gentle mixing.
16. Centrifuge the DNA solution at 16000 x g for 10 minutes at 4 °C. Transfer the aqueous portion to a clean tube without disturbing the pellet and store at 4 °C.
17. Add equal volume of 20% PEG precipitation buffer to the extracted DNA solution. Mix well by pipetting or inversion.
18. Incubate the PEG/DNA mixture for 15 minutes at 37 °C.
19. Centrifuge the PEG/DNA mixture for 15 minutes at approximately 15000 x g at room temperature.
20. Pour off supernatant or remove by pipetting. Wash the walls of the tube and DNA pellet with 80% ethanol (1.25 times volume of the original PEG/DNA mixture). Pour off ethanol or remove by pipetting.
21. Repeat wash once for a total of two washes (step 20).

22. Completely dry any residual ethanol by vacufuge at low heat (4-6 minutes).
23. Re-suspend the pellet in TE or H₂O using approximately equal volume as original DNA solution.
24. Centrifuge the re-suspended DNA solution at 15000 x g for 15 minutes.
25. Transfer DNA solution to a clean tube without disturbing the pellet.

4. Testing of the DNA extraction method by the Community Reference Laboratory for GM Food and Feed

The aim of the experimental testing was to verify that the DNA extraction method provides DNA of suitable quantity and quality for the intended purpose. The DNA extraction method should allow preparation of the analyte in quality and quantity appropriate for the analytical method used to quantify the event-specific analyte versus the reference analyte.

The CRL-GMFF tested the method proposed by the applicant on samples of food and feed consisting of ground soybean seeds provided by the applicant.

To assess the suitability of the DNA extraction method for real-time PCR analysis, the extracted DNA was tested using a qualitative PCR run on the real-time PCR equipment.

4.1 Preparation of samples

About 200 g of soybean seed material were ground using a GRINDOMIX GM 200 (Retsch GmbH) mixer.

4.2 DNA extraction

DNA was extracted following the method described above and provided by the applicant; the DNA extraction was carried out on 6 replicated test portions.

4.3 DNA concentration, yield and repeatability

Concentration of the DNA extracted was determined by fluorescence detection using the PicoGreen dsDNA Quantitation Kit (Molecular Probes).

Each DNA extract was measured twice, and the two values were averaged. DNA concentration was determined on the basis of a five point standard curve ranging from 1 to 500 ng/μl using a Biorad VersaFluor fluorometer.

The DNA concentration for all samples is reported in the Table 1 below.

Table 1. DNA concentration (ng/ μ l) of six extractions from samples of soybean seeds.

Sample	Concentration (ng/ μ l)
1	480.4
2	361.2
3	444.6
4	478.0
5	439.9
6	434.9

✓ DNA concentration (ng/ μ l)

Overall average of all samples: 439.8 ng/ μ l
 Standard deviation of all samples 43.2 ng/ μ l
 Coefficient of variation 9.8 %


✓ Yield (total volume of DNA solution: 1000 μ l)

Overall average of all samples: 440 μ g
 Standard deviation 43.2 μ g
 Coefficient of variation 9.8 %

4.4 Fragmentation of DNA

The size of the extracted DNA was evaluated by agarose gel electrophoresis; 6 μ l of a 1:10 dilution of the DNA solution were analysed on a 1.0% agarose gel (Figure 1).

Figure 1. Agarose gel electrophoresis of 6 genomic DNA samples extracted from soya seeds, lanes 1-6; M: Lambda DNA/EcoRI+HindIII Marker.



The six genomic DNA samples extracted as described above appeared as distinct fluorescent banding patterns migrating through the gel corresponding to high molecular weight DNA. None of the DNA samples showed indications of significant degradation ('smearing').

4.5 Purity / Absence of PCR inhibitors

In order to assess the purity and to confirm the absence of PCR inhibitors, the extracted DNA solutions were adjusted to a concentration of 50 ng/μl (hereafter referred as "undiluted" samples). Subsequently, fourfold serial dilutions of each extract were prepared with 0.2x TE buffer (1:4, 1:16, 1:64, 1:256) and analysed using a real-time PCR system detecting the target sequence of the endogenous control gene lectin, *lec*. The Ct values obtained for "undiluted" and diluted DNA samples are reported in Table 2.

Table 2. Ct values of undiluted and fourfold serially diluted DNA extracts after amplification of soybean lectin gene, *lec*.

DNA extract	Undiluted (50 ng/μl)	Diluted			
		1:4	1:16	1:64	1:256
1	21.05	22.91	25.42	27.23	28.74
2	21.07	23.26	25.23	27.04	29.58
3	20.98	22.95	24.79	26.87	28.86
4	20.92	22.79	24.65	26.73	28.94
5	21.07	22.80	24.82	26.74	28.66
6	21.15	22.86	25.01	26.82	28.65

Table 3 below reports the comparison of extrapolated Ct values versus measured Ct values for all samples and the values of linearity (R^2) and slope of all measurements.

To measure inhibition, the Ct values of the four diluted samples were plotted against the logarithm of the dilution and the Ct value for the "undiluted" sample (50 ng/μl) was extrapolated from the equation calculated by linear regression.

Subsequently, the extrapolated Ct for the "undiluted" sample was compared with the measured Ct. The evaluation is carried out considering that PCR inhibitors are present if the measured Ct value for the "undiluted" sample is suppressed by > 0.5 cycles from the calculated Ct value. In addition, the slope of the curve should be between -3.6 and -3.1.

Table 3. Comparison of extrapolated Ct values versus measured Ct values (amplification of soybean lectin gene, *lec*)

DNA extraction	R^2	Slope*	Ct extrapolated	mean Ct measured	ΔCt^{**}
1	0.9819	-3.207	21.25	21.05	0.20
2	0.9876	-3.451	21.08	21.07	0.02
3	0.9979	-3.290	20.92	20.98	0.06
4	0.9948	-3.411	20.64	20.92	0.28
5	0.9973	-3.238	20.88	21.07	0.19
6	0.9968	-3.185	21.04	21.15	0.10

*The expected slope for a PCR with 100% efficiency is -3.32

**delta Ct = abs (Ct extrapolated - Ct measured)

All ΔC_t values of extrapolated versus measured C_t are < 0.5 .

R^2 of linear regression is > 0.98 for all DNA samples. All the slopes of the curves are between -3.1 and -3.6, with no exceptions.

Note: Numerical values presented in the tables of this report were rounded keeping two digits for values ≤ 1 , one digit for values between 1 and 10 and no digit for values ≥ 10 . The calculations in the MS Excel files however were done over not rounded data. This approach might create generate small inconsistencies in the numerical values reported in the tables but it allows a higher precision in the final results.

5. Conclusion

The data reported confirm that the extraction method, applied to soybean seeds provided by the applicant, produces DNA of suitable quantity and quality for subsequent PCR based detection applications. The method is consequently applicable to samples of soybean seeds provided as samples of food and feed in accordance with the requirements of Annex I-2.C.2 to Commission Regulation (EC) No 641/2004.

6. Quality assurance

The CRL-GMFF carries out all operations according to ISO 9001:2000 (certificate number: CH-32232) and ISO 17025:2005 (certificate number: DAC-PL-0459-06-00) [DNA extraction, qualitative and quantitative PCR in the area of Biology (DNA extraction and PCR method validation for the detection and identification of GMOs in food and feed materials)]

7. References

1. Murray M.G and Thompson W.F., 1980. Rapid isolation of high molecular weight plant DNA. *Nucleic Acids Research* 8, 4321-4325.
2. Sambrook J. and D. W. Russell. *Molecular Cloning. A laboratory manual*. Third edition. 2001. Cold Spring Harbor Laboratory Press.

Page intentionally left blank

Europe Direct is a service to help you find answers to your questions about the European Union
Freephone number (*): 00 800 6 7 8 9 10 11

(*) Certain mobile telephone operators do not allow access to 00 800 numbers or these calls may be billed.

A great deal of additional information on the European Union is available on the Internet.
It can be accessed through the Europa server <http://europa.eu/>.

How to obtain EU publications

Our priced publications are available from EU Bookshop (<http://bookshop.europa.eu>),
where you can place an order with the sales agent of your choice.

The Publications Office has a worldwide network of sales agents.
You can obtain their contact details by sending a fax to (352) 29 29-42758.

European Commission
Joint Research Centre – Institute for Health and Consumer Protection

Title: Report on the Validation of a DNA Extraction Method for Soybean Seeds

Author(s): European Union Reference Laboratory for Genetically Modified Food and Feed

2014 – 11 pp. – 21.0 x 29.7 cm

Abstract

This report describes the validation of a DNA extraction protocol derived from the publicly available “CTAB” method (1) and its applicability on the samples of food and feed provided by the applicant. This protocol can be used for the extraction of DNA from soybean seeds.

The purpose of the DNA extraction method described is to provide DNA with purity suitable for real-time PCR based detection methods.

This protocol is recommended to be executed by skilled laboratory personnel as the procedures comprise the use of hazardous chemicals and materials. It is strongly advised to take particular notice of products safety recommendations and guidelines.

JRC Mission

As the Commission's in-house science service, the Joint Research Centre's mission is to provide EU policies with independent, evidence-based scientific and technical support throughout the whole policy cycle.

Working in close cooperation with policy Directorates-General, the JRC addresses key societal challenges while stimulating innovation through developing new methods, tools and standards, and sharing its know-how with the Member States, the scientific community and international partners.

Serving society
Stimulating innovation
Supporting legislation