



EUROPEAN COMMISSION
DIRECTORATE GENERAL JRC
JOINT RESEARCH CENTRE
INSTITUTE FOR HEALTH AND CONSUMER PROTECTION
COMMUNITY REFERENCE LABORATORY FOR GM FOOD AND FEED



Event-specific method for the quantitation of maize line MON 863 using real-time PCR

Protocol

Method development:

Monsanto Biotechnology Regulatory Sciences
(only for the PCR part)

Method validation:

Joint Research Centre – European Commission
Biotechnology & GMOs Unit

Contents

1. GENERAL INFORMATION AND SUMMARY OF THE METHODOLOGY.....	4
2. VALIDATION STATUS AND PERFORMANCE CHARACTERISTICS.....	5
2.1 GENERAL	5
2.2 COLLABORATIVE TRIAL.....	5
2.3 LIMIT OF DETECTION	5
2.4 LIMIT OF QUANTITATION.....	5
2.5 MOLECULAR SPECIFICITY.....	6
3. PROCEDURES.....	6
3.1 GENERAL INSTRUCTIONS AND PRECAUTIONS	6
3.2 DNA EXTRACTION.....	7
3.3 SPECTROPHOTOMETRIC MEASUREMENT OF DNA CONCENTRATION.....	8
3.3.1 Measurement of a reference DNA solution.....	8
3.3.2 Measurement of a test DNA solution of unknown concentration.....	8
3.3.3 Evaluation.....	8
3.4 REAL-TIME PCR FOR QUANTITATIVE ANALYSIS OF MON 863 MAIZE.....	9
3.4.1 General.....	9
3.4.2 Calibration.....	9
3.4.3 Real-time PCR set-up.....	10
3.5 DATA ANALYSIS	11
3.6 CALCULATION OF RESULTS	12
4. MATERIALS.....	12
4.1 EQUIPMENT (EQUIVALENTS MAY BE SUBSTITUTED).....	12
4.2 REAGENTS.....	13
4.3 PRIMERS AND PROBES	14
5. BUFFERS AND SOLUTIONS.....	14
6. REFERENCES	17

<u>Document Approval</u>		
Name / Function	Date	Signature
Marco Mazzara <i>Sector Head</i>	16/02/2005	Signed
Stephane Cordeil <i>Quality Manager</i>	16/02/2005	Signed
Guy Van den Eede <i>B&GMOs Unit Head</i>	16/02/2005	Signed

Address of contact laboratory:

European Commission, Joint Research Centre
Institute for Health and Consumer Protection (IHCP)
Biotechnology and GMOs Unit – Community Reference Laboratory
Via Fermi 1, 21020 Ispra (VA) - Italy

1. General information and summary of the methodology

This protocol describes an event-specific real-time quantitative TaqMan[®] PCR procedure for the determination of the relative content of event MON 863 DNA to total maize DNA in a sample. The procedure includes the following three modules:

- a) DNA extraction: CTAB DNA extraction and purification protocol
- b) Spectrophotometric quantitation of the amount of total DNA
- c) Quantitative real-time PCR methodology specific for the NK603 event

The PCR assay has been optimised for use in an ABI Prism[®] 7700 sequence detection system. Other systems may be used, but thermal cycling conditions must be verified. The use of 200 ng of template DNA per reaction well is recommended.

DNA is extracted by means of a CTAB DNA extraction and purification protocol. For references, see Murray and Thompson (1980), Wagner *et al.* (1987) and Zimmermann *et al.* (1998). The protocol has been validated for soybeans (Anon, 1998), potato (Anon, 1996) and tomato (Anon, 1999). It has been tested for maize in a multi-laboratory pre-validation. The method was adopted from: Foodstuffs – Methods of analysis for the detection of genetically modified organisms and derived products – Nucleic Acid Extraction. CEN/TC 275/WG11N0031. Draft November 2002.

Subsequently, purified DNA is quantified by means of spectrophotometry in order to determine the amount of DNA to be analysed by means of real-time PCR. The procedure “Basic ultraviolet spectrometric method” has been adopted from the Annex B “Methods for the quantification of the extracted DNA” of the prEN ISO 21571:2002. The method has been widely used and ring-tested (Anon. 2002).

For specific detection of event MON 863 genomic DNA, a 84-bp fragment of the region that spans the 5′ insert-to-plant junction in maize event MON 863 is amplified using two specific primers. PCR products are measured during each cycle (real-time) by means of a target-specific oligonucleotide probe labelled with two fluorescent dyes: FAM as a reporter dye at its 5′ end and TAMRA as a quencher dye at its 3′ end.

For relative quantitation of event MON 863 DNA, a maize-specific reference system amplifies a 70-bp fragment of *adh1*, a maize endogenous gene, using a pair of *adh1* gene-specific primers and an *adh1* gene-specific probe labelled with FAM and TAMRA as described above.

The measured fluorescence signal passes a threshold value after a certain number of cycles. This threshold cycle is called the “Ct” value. For quantitation of the amount of

event MON 863 DNA in a test sample, event MON 863 and *adh1* Ct values are determined for the sample. Standard curves are then used to calculate the relative content of event MON 863 DNA to total maize DNA.

2. Validation status and performance characteristics

2.1 General

The method has been optimised for maize seeds, grain and flour containing mixtures of genetically modified MON 863 and conventional maize.

The reproducibility and trueness of the method was tested through collaborative trial using samples of the candidate CRM IRMM-416 series.

2.2 Collaborative trial

The method was validated in a collaborative trial by the Joint Research Centre (JRC) of the European Commission. The study was undertaken with 12 laboratories.

Each participant received ten unknown samples. The samples consist of five reference materials (candidate CRM IRMM-416) of dried maize powder containing mixtures of genetically modified MON 863 maize in conventional maize (w/w) between 0.0 % and 10%.

For each unknown sample one DNA extraction has been carried out. Each test sample was analyzed by PCR in four repetitions. The study was designed as a blind duplicate collaborative trial. Each laboratory received each level of GM MON 863 in two unknown samples, and the two replicates for each GM level were analyzed in two PCR plates.

A detailed validation report can be found under <http://gmo-crl.jrc.it/statusofdoss.htm>

2.3 Limit of detection

According the method developer, the relative LOD of the method is at least 0.05%. The relative LOD was not assessed in a collaborative trail. The lowest relative concentration of the target sequence included in collaborative trail was 0.1%.

2.4 Limit of quantitation

According the method developer, the relative LOQ of the method is 0.1%. The lowest relative concentration of the target sequence included in collaborative trail was 0.1%.

2.5 Molecular specificity

The method utilizes the unique DNA sequence at the junction of the insert and the genomic DNA flanking the insert. The sequence is specific to MON 863 and thus imparts specificity to the detection method.

The specificity was experimentally tested against DNA extracted from plant materials containing the specific targets of GA21, NK603, MON810 maize, and from conventional corn, Roundup Ready® soybean, conventional soybean, Roundup Ready® canola, conventional canola and Roundup Ready® wheat. None of the materials yielded detectable amplification.

The target sequence is a single copy sequence in the haploid MON 863 genome.

3. Procedures

3.1 General instructions and precautions

- The procedures require experience of working under sterile conditions.
- Maintain strictly separate working areas for DNA extraction, PCR set-up and amplification.
- All the equipment used must be sterilized prior to use and any residue of DNA has to be removed.
- In order to avoid contamination, filter pipette tips protected against aerosol should be used.
- Use only powder-free gloves and change them frequently.
- Clean lab-benches and equipment periodically with 10% sodium hypochloride solution (bleach).
- Pipettes should be checked regularly for precision and calibrated, if necessary.

3.2 DNA extraction

- a. Moisten 200 mg of sample with 300 µl of sterile deionised water in a 1.5 ml tube.
- b. Mix with a sterile loop until homogeneity is reached.
- c. Add 700 µl of CTAB-buffer pre-warmed to 65°C; mix with a loop or a clean spatula
- d. Add 10 µl of RNase solution; shake.
- e. Incubate at 65° C for 30 min.
- f. Add 10 µl of Proteinase K solution; mix smoothly.
- g. Incubate at 65° C for 30 min.
- h. Centrifuge for 10 min at 12000 g
- i. Transfer supernatant to a 1.5 ml tube, containing 500 µl chloroform; shake for 30 sec.
- j. Centrifuge for 15 min at 12000 g until phase separation occurs.
- k. Transfer the aqueous upper phase into a new 1.5 ml tube containing 500 µl chloroform; shake.
- l. Centrifuge for 5 min at 12000 g.
- m. Transfer upper layer to a new 1.5 ml tube.
- n. Add 2 volumes of CTAB precipitation solution, mix by pipetting.
- o. Incubate for 60 min at room temperature.
- p. Centrifuge for 5 min at 13000 rpm; discard the supernatant.
- q. Dissolve precipitate in 350 µl NaCl (1.2 M).
- r. Add 350 µl chloroform and shake for 30 sec.
- s. Centrifuge for 10 min at 12000 g until phase separation occurs.
- t. Transfer upper layer to a new reaction tube.
- u. Add 0.6 volumes of isopropanol, mix smoothly by inversion. Incubate for 20 min at room temperature.
- v. Centrifuge for 10 min at 12000 g. Discard the supernatant.
- w. Add 500 µl of 70% ethanol solution and shake carefully.
- x. Centrifuge for 10 min at 12000 g. Discard the supernatant.
- ATTENTION:** drain the supernatant carefully. DNA pellets may detach from the bottom of the tube at this stage.
- y. Dry pellets and re-dissolve DNA in 100 µl sterile, TE buffer.
- z. **IMPORTANT:** for thorough homogenisation of the DNA solution, it is recommended to re-suspend the sample by gentle agitation at +4°C for approximately 12 h.

The DNA solution may be stored at ~ 4°C for a maximum of one week, or at -20°C for long-term storage.

3.3 Spectrophotometric measurement of DNA concentration

3.3.1 Measurement of a reference DNA solution

The correct calibration of the spectrometer can be verified as follows, with the use of a reference DNA solution:

- a) For blank measurement only dilution buffer is used to fill the measurement vessel.
- b) The reference DNA solution (Calf Thymus or Herring Testes DNA or Lambda DNA) is filled into the measurement vessel.

Absorption is measured for both blank and reference DNA solutions at $\lambda = 260$ nm and $\lambda = 320$ nm.

3.3.2 Measurement of a test DNA solution of unknown concentration

- a) Blank measurement: mix the dilution buffer with a 2M sodium hydroxide solution, at the final NaOH concentration of 0.2M. This solution is used for the blank measurement.
- b) Mix the DNA solutions with a 2M sodium hydroxide solution and, if needed, with dilution buffer, at the final NaOH concentration of 0.2M.
- c) Measure the absorption after 1 min incubation time for both blank and reference DNA solution at $\lambda = 260$ nm and $\lambda = 320$ nm. The reading is stable for at least 1 h.

Example for blank measurement: Mix 90 μ l dilution buffer and 10 μ l of 2M sodium hydroxide solution and transfer to a 100 μ l measurement vessel.

Example for the test DNA solution: Mix 80 μ l of dilution buffer or water, 10 μ l of 2M sodium hydroxide solution, 10 μ l of DNA solution of unknown concentration and transfer to a 100 μ l measurement vessel.

3.3.3 Evaluation

The absorption (OD) at 320 nm (background) is subtracted from the absorption at 260 nm resulting in the corrected absorption at 260 nm. If the corrected OD at 260 nm equals to 1, then the estimated DNA concentration is 38 μ g/ml for single stranded DNA (denatured with sodium hydroxide).

Reliable measurements require OD values at $\lambda=260$ nm greater than 0.05.

The concentration of the double stranded test DNA solution is finally calculated taking into consideration the denaturation and the dilution factor applied.

3.4 Real-time PCR for quantitative analysis of MON 863 maize

3.4.1 General

The PCR set-up for the taxon specific target sequence (*Adh1*) and for the GMO (MON 863) target sequence should be carried out in separate vials. Multiplex PCR (using differential fluorescent labels for the probes) has not been tested or validated.

The use of 200 ng of template DNA per reaction well is recommended.

The method is developed for a total volume of 50 μ l per reaction mixture with the reagents as listed in Table 1.

3.4.2 Calibration

Separate calibration curves with each primer/probe system are generated in the same analytical amplification run.

The calibration curves consist of five dilutions of DNA extracted from the 10% candidate CRM IRMM-416-3. A series of one to four dilution intervals at a starting concentration of 102,752 maize genome copies may be used (corresponding to 280 ng of DNA with one maize genome assumed to correlate to 2.725 pg of haploid maize genomic DNA) (Arumuganathan & Earle, 1991).

A calibration curve is produced by plotting Ct-values against the logarithm of the target copy number for the calibration points. This can be done e.g. by use of spreadsheet software, e.g. Microsoft Excel, or directly by options available with the sequence detection system software.

The copy numbers measured for the unknown sample DNA is obtained by interpolation from the standard curves.

3.4.3 Real-time PCR set-up

1. Thaw, mix gently and centrifuge the required amount of components needed for the run. **Keep thawed reagents at 1-4°C on ice.**
2. In two reaction tubes (one for MON 863 system and one for the *adh1* system) on ice, add the following components (Table 1) in the order mentioned below (except DNA) to prepare the master mixes.

Table 1. Amplification reaction mixtures in the final volume/concentration per reaction well, for MON 863/*adh1* specific systems.

Component	Final concentration	µl/reaction
TaqMan® Universal PCR Master Mix (2X)	1x	25 µl
Primer MON863-F/ <i>adh1</i> -F	150 nM	-
Primer MON863-R/ <i>adh1</i> -R	150 nM	-
Probe MON863/ <i>adh1</i>	50 nM	-
Nuclease free water		up to 50 µl
Template DNA (maximum 280 ng, see 3.4.1 and 3.4.2)		5 µl
Total reaction volume:		50 µl

3. Mix gently and centrifuge briefly.
4. Prepare two 1.5 ml reaction tubes (one for the NK603 and one for the *adh1* master mix) for each DNA sample to be tested (standard curve samples, unknown samples and control samples).
5. Add to each reaction tube the correct amount of master mix (e.g. 45 x 3 = 135 µl master mix for three PCR repetitions). Add to each tube the correct amount of DNA (e.g. 5 x 3 = 15 µl DNA for three PCR repetitions). Low-speed vortex each tubes at least three times for approx 30 sec. This step is mandatory to reduce the variability among the repetitions of each sample to a minimum.
6. Spin down the tubes in a micro-centrifuge. Aliquot 50 µl in each well. Seal the reaction plate with optical cover or optical caps. Centrifuge the plate at low speed (e.g. approximately 250 x *g* for 1 minute at 4 °C to room temperature) to spin down the reaction mixture.

7. Place the plate into the instrument.
8. Run the PCR with cycling conditions described in Table 2:

Table 2. Reaction conditions.

Step	Stage	T°C	Time (sec)	Acquisition	Cycles
1	UNG pre-PCR decontamination	50 °C	120"	No	1x
2	Activation of DNA polymerase and denaturation	95 °C	600"	No	1x
3	Denaturation	95 °C	15"	No	45x
4	Amplification Annealing & Extension	60 °C	60"	Measure	

3.5 Data analysis

Subsequent to the real-time PCR, analyse the run following the procedure below:

a) Set the threshold: display the amplification curves of one system (e.g. *adh1*) in logarithmic mode. Locate the threshold line in the area where the amplification profiles are parallel (exponential phase of PCR) and where there is no "fork effect" between repetitions of the same sample. Press the update button to ensure changes affect Ct values. Switch to the linear view mode by clicking on the Y axis of the amplification plot, and check that the threshold previously set falls within the geometric phase of the curves.

b) Set the baseline: determine the cycle number at which the threshold line crosses the first amplification curve and set the baseline three cycles before that value (e.g. earliest Ct = 25, set the baseline crossing at Ct = 25 – 3 = 22).

c) Save the settings

d) Repeat the procedure described in a) and b) on the amplification plots of the other system (e.g. MON 863 system).

e) Save the settings and export all the data into an Excel file for further calculations.

3.6 Calculation of results

After having defined a threshold value within the logarithmic phase of amplification as described above, the instruments software calculated the Ct-values for each reaction.

The standard curves are generated both for the *adh1* and MON 863 specific system by plotting the Ct-values measured for the calibration points against the logarithm of the DNA copy numbers, and by fitting a linear regression line into these data.

Thereafter, the standard curves are used to estimate the copy numbers in the unknown sample DNA by interpolation from the standard curves.

For the determination of the amount of MON 863 DNA in the unknown sample, the MON 863 copy number is divided by the copy number of the maize reference gene (*adh1*) and multiplied by 100 to obtain the percentage value ($GM\% = \text{MON 863}/adh1 * 100$).

4. Materials

4.1 Equipment (equivalents may be substituted)

DNA extraction:

- Water bath or heating block
- Microcentrifuge
- Micropipettes
- Vortexer
- 1.5/2.0 ml tubes
- Tips and filter tips for micropipettes
- Rack for reaction tubes
- Vinyl or latex gloves
- Optional: vacuum dryer apt to dry DNA pellets

Spectrophotometry:

- UV spectrophotometer. Single beam, double beam or photodiode array instruments are suitable.
- Vortexer
- Measurement vessels. e.g. quartz cuvettes or plastic cuvettes suitable for UV detection at a wavelength of 260 nm. The size of the measurement vessels used determines the volume for measurement. This should be one of the following: half micro cuvettes (1000 μl), micro cuvettes (400 μl), ultra micro cuvettes (100 μl)

and quartz capillaries (3 μ l to 5 μ l). The optical path of standard cuvettes is usually 1 cm.

Real-time PCR:

- ABI Prism[®] 7700 Sequence Detection System. Applied Biosystems Part No 7700-01-200/208.
- ABI Prism[®] 7900HT Sequence Detection System. Applied Biosystems Part No 4329002 or 4329004.
- Software: Sequence Detection System version 1.7 (Applied Biosystems Part No 4311876) or equivalent versions.
- MicroAmp[®] optical 96-Well reaction plates. Applied Biosystems Part No N801-0560).
- MicroAmp[®] optical adhesive covers. Applied Biosystems Part No 4311971.
- MicroAmp Optical caps. Applied Biosystems Part No. No 801-0935.
- Microcentrifuge
- Micropipettes
- Vortex
- Rack for reaction tubes
- 1.5/2.0 ml tubes

4.2 Reagents

DNA extraction:

- CTAB: Cetyltrimethylammonium Bromide (Ultrapure grade)
- TRIS: Tris[hydroxymethyl] aminomethane hydrochloride (Molecular Biology grade)
- EDTA: Ethylenediaminetetraacetic acid, disodium salt (titration 99.9%)
- Ethanol (96% at least)
- Isopropanol (99.7% at least)
- Chloroform (99% at least)
- NaCl (99% at least)
- NaOH (98% at least, anhydrous)
- Distilled sterile water
- RNase A solution 10 mg/ml
- Proteinase K solution 20 mg/ml

Spectrophotometry:

- Dilution buffer: TRIS: Tris[hydroxymethyl] aminomethane hydrochloride (Molecular Biology grade). 10 mM, pH 9.0.
- NaOH (98% at least, anhydrous)
- Hydrochloric acid (HCl), ϕ (HCl) = 37 %

- Herring Testes DNA, Calf Thymus DNA, or Lambda DNA

Real-time PCR:

- TaqMan® Universal PCR Master Mix (2X). Applied Biosystems Part No 4304437

4.3 Primers and Probes

Name	Oligonucleotide DNA Sequence (5' to 3')
<i>GMO target sequence</i>	
MON863 primer F	GTAGGATCGGAAAGCTTGGTAC
MON863 primer R	TGTTACGGCCTAAATGCTGAACT
MON863 probe	6-FAM-TGAACACCCATCCGAACAAGTAGGGTCA-TAMRA
<i>Reference gene target sequence</i>	
<i>Adh1</i> primer F	CCAGCCTCATGGCCAAAG
<i>Adh1</i> primer R	CCTTCTTGCGGCTTATCTG
<i>Adh1</i> probe	6-FAM-CTTAGGGGCAGACTCCCGTGTTCCCT-TAMRA

5. Buffers and Solutions

The following describes the preparation, storage and stability of the buffers used in this procedure. Volume may be scaled as needed. Equivalent reagent may be substituted.

DNA extraction:

- **CTAB buffer (1 litre)**

Weight and mix in an appropriate cylinder:

20 g/l CTAB	20 g
1.4 M NaCl	82 g
0.1 M Tris-HCl	15.75 g
20 mM Na ₂ EDTA	7.5 g

- Add 500 ml of sterile distilled water.
- Adjust pH to a value of 8.0 with 1M NaOH.
- Fill up to 1000 ml and autoclave.

Store at 4° C for up to 6 months.

- **CTAB-precipitation solution (200 ml)**

Weight and mix in an appropriate cylinder:

5 g/l CTAB	1 g
0.04 M NaCl	0.5 g

- Add 100 ml of distilled water.
- Adjust pH to a value of 8.0 with 1 M NaOH.
- Fill up to 200 ml and autoclave.

Store at 4° C for up to 6 months.

- **NaCl 1.2 M (100 ml)**

- Dissolve 7 g of NaCl in 100 ml sterile distilled water in a cylinder.
- Autoclave

Store at room temperature for up to 5 years

- **Ethanol-solution ~70 % (v/v) (100 ml)**

- Mix 70 ml of pure ethanol with 30 ml of sterile distilled water

Store at room temperature or at -20° C for up to 5 years

- **NaOH 1M (50 ml)**

- Dissolve 2 g of NaOH in 50 ml of sterile water in a cylinder or a 50 ml conical tube.

Store at room temperature for up to 6 months

- **TE buffer, pH 7.0 (Tris/HCl 10 mM, EDTA 1 mM, pH 7.0) (250 ml)**

- Mix 100 ml of nuclease-free water, 2.5 ml of 1M Tris, pH 8.0 and 0.5 ml of 0.5M EDTA
- Adjust pH to 7.0 with HCl
- Adjust final volume to 250 ml with nuclease-free water
- Filter sterilise

Store at room temperature for up to 5 years

- **RNase A 10 mg/ml**

- Dissolve the RNase A at a final concentration of 10 mg/ml in sterile water.
- If indicated by supplier: boil the RNase A solution at 95°C for 15' to remove any residual nuclease activity.
- Aliquot solution as appropriate (thawing and re-freezing should be avoided)

Store aliquots at -20° C for up to 6 months

- **Proteinase K 20 mg/ml**

- a. Dissolve the Proteinase K at a final concentration of 20 mg/ml in sterile distilled water according to the supplier specifications.
- b. Aliquot solution as appropriate (thawing and re-freezing should be avoided)

Store aliquots at -20° C for up to 6 months

Spectrophotometry:

- **Reference DNA solution**

A DNA 10 mg/ml stock solution is prepared by dissolving 100 mg DNA (from Herring Testes or from Calf Thymus or Lambda DNA) in 10 ml dilution buffer (TRIS/HCl 10 mM, pH 9.0). At this concentrations DNA dissolves and homogenises slowly and the resulting solution is very viscous. The stock solution is further diluted with dilution buffer up to the desired working concentration (e.g. 25 µg/ml).

- **NaOH 2M (50 ml)**

- a. Dissolve 4 g of NaOH in 50 ml of sterile water in a cylinder or a 50 ml conical tube.

Store at room temperature for up to 5 years

6. References

- Anon. 1996. Untersuchung von Lebensmitteln: Nachweis einer gentechnischen Veränderung von Kartoffeln durch Amplifizierung der veränderten DNA-Sequenz mit Hilfe der PCR (Polymerase Chain Reaction) und Hybridisierung des PCR-Produktes mit einer DNA-Sonde. L 24.01-01, February 1996, revised in January 1997.. In: Amtliche Sammlung von Untersuchungsverfahren nach § 35 LMBG; Verfahren zur Probenahme und Untersuchung von Lebensmitteln, Tabakerzeugnissen, kosmetischen Mitteln und Bedarfsgegenständen/ BgVV. Loseblattausgabe, Stand Jan 1997 Bd.1. Berlin, Köln. Beuth Verlag GmbH.
- Anon 1998. Untersuchung von Lebensmitteln: Nachweis einer gentechnischen Veränderung von Sojabohnen durch Amplifizierung der veränderten DNA-Sequenz mit Hilfe der PCR (Polymerase Chain Reaction) und Hybridisierung des PCR-Produktes mit einer DNA-Sonde. L 23.01.22-1, March 1998. In: Amtliche Sammlung von Untersuchungsverfahren nach § 35 LMBG; Verfahren zur Probenahme und Untersuchung von Lebensmitteln, Tabakerzeugnissen, kosmetischen Mitteln und Bedarfsgegenständen/ BgVV. Loseblattausgabe, Stand March 1998 Bd.1. Berlin, Köln. Beuth Verlag GmbH.
- Anon. 1999. Untersuchung von Lebensmitteln: Nachweis einer gentechnischen Veränderung von Tomaten durch Amplifizierung der veränderten DNA-Sequenz mit Hilfe der PCR (Polymerase Chain Reaction) und Hybridisierung des PCR-Produktes mit einer DNA-Sonde L 25.03.01-1, November 1999. In: Amtliche Sammlung von Untersuchungsverfahren nach § 35 LMBG; Verfahren zur Probenahme und Untersuchung von Lebensmitteln, Tabakerzeugnissen, kosmetischen Mitteln und Bedarfsgegenständen/ BgVV. Loseblattausgabe, Stand Nov 1999 Bd.1. Berlin, Köln. Beuth Verlag GmbH.
- Anon. 2002. Swiss food manual, Chapter 52B, Section 1 to 5. Eidgenössische Drucksachen und Materialzentrale, CH-5005 Bern. Available on CD.
- Arumuganathan, K., Earle, E.D. (1991). Nuclear content of some important plant species. *Plant Mol Biol Reporter* 9, 208-218.
- Murray, M.G. and Thompson, W.F. (1980). Rapid isolation of high molecular weight plant DNA. *Nucleic Acids Research* 8, 4321–4325.
- Wagner, D.B., Furnier, G.R., Saghay-Marooof, M.A., Williams, S.M., Dancik, B.P. and Allard, R.W. (1987). Chloroplast DNA polymorphisms in lodgepole and jack pines and their hybrids. *Proceedings of the National Academy of Science USA* 84, 2097–2100.
- Zimmermann, A., Lüthy, J. and Pauli, U. (1998). Quantitative and qualitative evaluation of nine different extraction methods for nucleic acids on soya bean food samples. *Zeitschrift für Lebensmittel-Untersuchung und -Forschung A* 207, 81–90.